

SUPPLY , DELIVERY AND INSTALLATION OF BIOMEDICAL TEST EQUIPMENT - 2025–
UVA PROVINCIAL DEPARTMENT OF HEALTH SERVICES

Prebid Meeting – 2025/10/03

- **CDDA/ NMRA certificate should be removed from the ITB 11.1 (e) and ITB 35.4 (d) of the bidding document for all packages**

PDHS/BMEU/TE/07/2025 – Infusion Device Analyzer

1. Clause No.02 (v) -

The clause “Bolus Volume Range: 0.01 – 99,999mL” is compulsory

- **The compulsory should be removed from the clause**

2. Clause No.02 (viii) -

The clause “Volume Measurement Range: 0 – 99999 mL” is compulsory

- **The compulsory should be removed from the clause**

PDHS/BMEU/TE/02/2025 – Electrical Safety Analyzer

1. Clause No.02 (xvi) -

The clause “ECG simulation , performance waveforms and respiration rate”

- **The respiratory rate should be removed from the clause**

2. Clause No.08 (i) -

The clause “Apnea 0-100 BrPM in 10 BrPM steps with normal waveforms”

- **The clause should be removed**

R. H. M. DASANATAKA
Financial Assistant
Department of Health Services - Uva
Tel : 055-4938038